



1. Plug in the Thermocouple NiCr-Ni, Type K with Miniature Plug into the destined socket at the measurement device.
2. Read out the measured value of the measurement device.
3. Measuring range of Mineral Insulated Thermocouple -50 °C...+260 °C (-58 °F...+500 °F)

Artikel	Artikelnummer
Air probe NiCr-Ni, teflon insulated, nominal length 1000 mm	0600 1301
Air probe NiCr-Ni, teflon insulated, nominal length 5000 mm	0600 1305

Operating Instruction

Air probe